

# SF6040

Laser diode driver

# Datasheet & User Manual

Before powering on your driver, read this manual thoroughly.

If you have any doubt or suggestion, please do not hesitate to contact us!

Maiman Electronics LLC, st. Sedova, 37, lit. "A", Saint-Petersburg, Russia, 192148

e-mail: info@maimanelectronics.com web site: www.maimanelectronics.com

#### 1. Features

- Low current ripple
- No need to adjust voltage
- Soft-start
- Adjustable overcurrent limit
- Thermal warning and shutdown
- Reverse current protection
- Crowbar circuit protection
- Universal controls
- NTC thermistor input

### 2. Applications

- Supplying laser diodes, bars and arrays
- Supplying high power LED arrays

#### 3. Description

The SF6040 is a non-isolated DC/DC SMPS (POL) with constant current output. Driver produces high stability and low ripple current. Protection features include an over current protection with adjustable limit, soft-start, thermal warning and shutdown, reverse current protection and crowbar circuit protection. This means the SF6040 can safely drive nonlinear and current quality sensitive loads like laser diodes and LEDs.

The driver can be controlled by analogue or digital signals.

SF6040 is available in thermal enhanced quarter-brick package with aluminum base plate to aid thermal dissipation. Driver can be mounted on any thermal conductive surface enough to dissipate driver losses.

## 4. Absolute maximum ratings\*

	MIN	MAX	UNIT
Vin+ to Vin-	-0.3	60	V
Operating temperature	-40	50	°C
Analogue control pins to GND	-0.3	5.5	V
RS-232 RXD to GND	-25	25	V
RS-232 TXD to GND	-13.2	13.2	V
UART RXD to GND	-0.3	5.5	V
UART TXD to GND	-0.3	5.5	V

<sup>\*</sup>Stresses beyond those listed under absolute maximum ratings may cause permanent damage to the device. These are stress ratings only, which do not imply functional operation of the device at these or any other conditions beyond those indicated under recommended operating conditions. Exposure to absolute-maximum-rated conditions for extended periods may affect device reliability.

# 5. Recommended operating conditions

	MIN	MAX	UNIT
Input voltage (V <sub>in</sub> )	12	55	V
Operating temperature	-10	40	°C
Analogue control pins to GND	0	5	V
RS-232 RXD to GND	-12	12	V
RS-232 TXD to GND	-12	12	V
UART RXD to GND	0	5	V
UART TXD to GND	0	5	V

# 6. Power supply requirements


The driver requires a DC power supply. The power supply must be able to cover the driver output power and losses. The power supply voltage must be at least 27% higher than the voltage drop of your laser diode.

#### 7. Electrical characteristics

PARAMETER	TEST CONDITIONS	MIN	TYP	MAX	UNIT
OUTPUT					
Output voltage (V <sub>out</sub> )		5		0.73*V <sub>in</sub>	٧
Output current (I <sub>out</sub> )		0		10	Α
Current ripple			15	25	mA
Pulse rate (F)	QCW-mode	0.1		100	Hz
Pulse duration	Set by RS-232 or UART	2	1/F - 2	5000	ms
	Set by analogue pin	0.5		1/F	
Rise time (Soft-start time)	I <sub>out</sub> = 10A	60	70	100	us
	$I_{out} = 5A$	60	80	100	us
Fall-time (Stop time)		20	60	80	us
Output capacitance			13.2		uF
INPUT					
Vin quiescent current	V <sub>in</sub> = 48V, no AUX load	10	20	30	mA
Input capacitance			17.6		uF
CONTROLS					
Enable pin low threshold				1	V
Enable pin high threshold		2.3			V
Interlock pin threshold				1	V
Interlock pin pull-up resistance			10		kΩ
Current set pin voltage vs. output			2		A/V
current			_		,,,,
Current set step	Set by RS-232 or UART		0.01		Α
Current set accuracy	1A < I <sub>out</sub> < 7A		+/- 5		%
•	7A < I <sub>out</sub> < 10A		+/- 1		%
Current set calibration		- 5		+ 5	%
AUX SUPPLY					
+/- 15V accuracy			+/- 2		%
+/- 15V output current			20	30	mA
5V accuracy			+/- 1.5		%
5V output current			200	300	mA
INTERNAL MEASURMENTS					
Internal measurements accuracy			+/-2		%
Current monitor pin voltage vs.			2		A/V
output current			_		, -
Voltage monitor pin voltage vs.			5		V/V
output voltage			-		,
External sensor temperature	NTC 10k	-10		150	°C
POWER DISSIPATION					
Driver losses	V <sub>in</sub> = 55V, I <sub>out</sub> = 10A	10	13	15	W

PROTECTIONS			
Crowbar clamp resistance	7.4	9.4	mΩ
Crowbar clamp response time	100	200	ns
Over-temperature warning threshold temp	60		°C
Over-temperature warning hysteresis	2		°C
Over-temperature shut down threshold temp	80		°C
Over-temperature shut down hysteresis	22		°C


# 8. Typical Performance Characteristics




98,00 96,00 95,00 94,00 91,00 91,00 91,00 90,00 1 2 3 4 5 6 7 8 9 10 Output current, A

Figure 1 – Power dissipation vs. output current at full load

Figure 2 – Efficiency vs. output current at full load








Figure 4 – Typical stop sequence

#### 9. Pin and terminal functions


Please, note polarity!

Never ground any lead of the output,
this may cause permanent damage to the laser diode and the driver!


Never use any grounded probes (e.g. from the oscilloscope) at the output!

Control pins are not isolated!

#### **Terminals**




#### **Controls**



# **Digital control connector**

Wurth WR-MM 6901 5700 08 72 or TE Connectivity 215083-8



Pin	Description
1	UART RXD
2	RS-232 RXD
3	UART TXD
4	RS-232 TXD
5	+5V auxiliary (duplicates pin 4 of analogue control connector)
6	Interlock (duplicates pin 5 of analogue control connector)
7	Crowbar status (duplicates pin 3 of analogue control connector)
8	GND (connected to Vin- terminal)

# **Analogue control connector**

Wurth WR-MM 6901 5700 14 72 or TE Connectivity 1-215083-4

PIN	I/O	Name	Description
1	ı	Enable	High = operates; Low = stop.
			Internally pulled down. TTL or CMOS signals only.
2		GND	Connected to Vin- terminal.
3	0	Crowbar status	High = fault; Low = normal operation. Internally pulled down.
4		+5V	Auxiliary +5V power supply.
•		134	Up to 200mA output current capability.
5	1	Interlock	Open = locked; Low = operates.
	•		Internally pulled up.
6	0	Over-temperature	High = $t^{\circ} > 60^{\circ}$ C; Low = $t^{\circ} < 58^{\circ}$ C.
		warning	Internally pulled down.
7		GND	Connected to Vin- terminal.
8		-15V	Auxiliary -15V power supply.
			Up to 20mA output current capability.
9	0	Voltage monitor	0-8V = 0-40V at the output.
10		+15V	Auxiliary +15V power supply.
			Up to 20mA output current capability.
11	0	Current monitor	0-5V = 0-10A at the output.
12		GND	Connected to Vin- terminal.
13	l	Current set	0-5V = 0-10A at the output.
14		NTC Interlock	Connect NTC thermistor 10k between this pin and GND.

#### 10. Analogue control description

#### **Enable**

The enable pin is active high 3.3V and 5V compatible logic input. Driver operates when enable pin is high. To ensure correct start-up without overshoot, the soft-start sequence initiates when the enable pin becomes high.

The enable pin can be used for QCW operation. The output may be pulsed by applying a TTL square signal to enable pin.

#### **Crowbar status**

Crowbar status pin reports the status of the protection shorting crowbar clamp across the output. The crowbar will short the output while over-current shutdown and over-temperature shutdown. While the protection shutdown the enable pin is ignored. Disconnect the driver from power to reset the protection status.

#### Interlock

The Interlock pin must be connected to GND in order for output current to be delivered. It can be used for external interlock e.g. door or over-temperature switches.

#### Over temperature warning

When the temperature of driver PCB exceeds 60°C the over-temperature warning will go high to indicate driver is in danger of shutting down due to over-temperature condition. When temperature of driver PCB exceeds 80°C, the driver will stop. The driver can be restarted when the temperature drops to 58°C.

#### **Voltage monitor**

The output voltage of the driver can be monitored by voltage monitor. 0-8V = 0-40V with +/-2% accuracy. The voltage is measured between LD+ and LD- terminals.

#### **Current monitor**

The output current of the driver can be monitored by current monitor. 0-5V = 0-10A with +/-2% accuracy.

#### **Current set**

The driver output current is set by applying an analog signal to the current set pin. 0-5V = 0-10A.

The current set pin can be used for analogue modulation by applying sign, square or ramp signal with the DC component. Analogue modulation amplitude depends on frequency. Please, control the output current while using this feature.

If you use arbitrary/function generator or lab PS for current set, make sure it is in High Z mode. Please, control the current set and current monitor pin voltages while getting started.

#### **NTC Interlock**

Thermistor connection. Allows you to measure the temperature of a laser or other devices. The measurement result is readable by the digital interface. Set upper and lower limits of the temperature at the outlet of which driver operation is blocked. When the temperature returns to the specified range, the driver operation resumes.

#### 11. How to get started

- 1. Unpack your driver;
- 2. Connect a dummy load (e.g. a Schottky diode) to the LD- and LD+ terminals. Please note polarity;
- 3. Connect the controls (analogue and/or digital);
- 4. Connect the power supply. Please note polarity;
- 5. Rotate the current limit potentiometer counterclockwise several times;
- 6. Turn on the power supply;
- 7. Set the current equal to the desired current limit threshold;
- 8. Enable the driver;
- 9. Rotate the potentiometer slowly while the crowbar status pin becomes high;
- 10. Restart the power supply to reset protection;
- 11. Make sure the over-current protection limit is set right;
- 12. Make sure that all controls are working;
- 13. Turn off the power supply and disconnect a dummy load;
- 14. Connect the laser diode to the LD- and LD+ terminals. Please note polarity.

#### 12. Cooling

The driver produces up to 15W of losses. Thus the base plate has to be mounted on a thermal conduction surface to ensure proper operation and prevent an over-temperature shutdown. If the conduction cooling is not enough an additional cooling may be achieved by placing the driver into the airflow of a fan.

#### 13. Internal protections

Several security features protect the laser diode and driver from damage. Reverse diode protects the laser diode from reverse currents. In case of an over-current or an over temperature condition, the control logic disables the driver and the output shorts by a crowbar clamp. A soft-start mechanism slowly raises the current at startup to protect the laser diode against current overshooting.

#### 14. Digital control description

When the input voltage applied the driver is always in "analogue current set, external enable and allowing interlock" state. Any other state should be set any time after powering the driver if needed.

#### **Default serial port settings:**

Baud rate	Data bits	Stop bits	Parity	Flow control
115200	8	1	none	none

Data exchange between the driver and the PC is only initiated by the PC. All commands are sent in plain text format. All commands are sent with prefix. Number of command follows the prefix without any symbols. If there is the value after the command they separates with "space" symbol. The command ends with "carriage return" symbol.

The format of the command to set the value (P-type);

Number of byte	Value	Comment
1	P (50h)	Set prefix
2-5	Number of the parameter	Hex-number of the parameter. For example, 0100h
6	'space' symbol (20h)	
7-10	New value of the parameter	Hex-value of the parameter. For example, 0000h
11	'return carriage' symbol <cr> (0Dh)</cr>	End of the command

The device does not respond to P-type commands by default. (see section "the protocol extension")

You can request the value of parameter by the J-type command. The device will return a value of requested parameter.

The format of the command to get the value (J-type)

Number of byte	Value	Comment
1	J (4Ah)	Request prefix
2-5	Number of the parameter	Hex-number of the parameter. For example, 0100h
6	'return carriage' symbol <cr> (0Dh)</cr>	End of the command

#### The format of the response

Number of byte	Value	Comment
1	K (4Bh)	Response prefix
2-5	Number of the parameter	Hex-number of the requested parameter
6	'space' symbol (20h)	
7-10	Returned value of the parameter	Hex-value of the parameter
11	'return carriage' symbol <cr> (0Dh)</cr>	End of the command

If the device could not recognize a command, it returns an error message with error code.

#### The format and codes of errors

Error (returned command)	Reasons (one or few)
E0000	1) Internal buffer of device is overflowed.
	2) Cannot find <cr> (0x0D) or\and <lf> (0x0A).</lf></cr>
	3) Format of command is invalid.
E0001	1) Unknown command (it does not P- or J-type command).
	2) The device failed to correctly interpret a command.
E0002	The CRC of received command (see section "the protocol extension").
K0000 0000	Request or set the parameter that does not exist.

# Available parameters and its description

Action			R/W	HEX- number of parameters
Francisco	Value		R/W	0100
Frequency	Minimum		R	0101
(0.1 Hz)	Maximum		R	0102
Dowation	Value		R/W	0200
Duration (0.1 ms)	Minimum		R	0201
(0.1 1118)	Maximum		R	0202
	Value		R/W	0300
Current	Minimum		R	0301
(0.01 A)	Maximum		R	0302
	Measured value (0.1A)		R	0307
Current set calibration (0.01%) <sup>1</sup>	Value		R/W	030E
Voltage (0.1 V)	Measured value		R	0407
State of the device (bit mask)	Start (Enable) Stop (Disable) Internal current set External current set External Enable Internal Enable Allow Interlock Deny Interlock Deny NTC Interlock Allow NTC Interlock	0008h 0010h 0020h 0040h 0200h 0400h 1000h 2000h 4000h 8000h	W	0700
Stateofthe device (bit mask)	0 bit  1 <sup>st</sup> bit  2 <sup>nd</sup> bit  4 <sup>th</sup> bit  6 <sup>th</sup> bit	1 – Device is powered on (always = 1)  0 – Stopped; 1 – Started  Current set: 0 – External; 1 – Internal  Enable: 0 – External; 1 – Internal  External NTC Interlock: 0 – Allowed; 1 – Denied  Interlock: 0 – Allowed; 1 – Denied	R	0700
Serial number	Return the hex-value of	f the serial number	R	0701

\_

 $<sup>^{1}</sup>$  Default – 100.00% (2710h), calibration range is from 95.00% (251Ch) to 105.00% (2904h).

Information about the purpose and possibilities of the device	Type of device 0-3 bits  5 – CW driver 6 – pulse driver			
	Appointment 4-7 bits 1 – Independent unit			
	Flags 8-11	8 – supports RS-232 9 – supports USB 10 – supports LAN 11 – supports TEC	R	0702
Information about	0 bit	1 – the device supports this option		0703
parameters that you can change	1 <sup>st</sup> bit	Frequency	R	
(bit mask)	2 <sup>nd</sup> bit	Duration		
(bit illusk)	3 <sup>rd</sup> bit	Current		
	0 bit Reserve			
Lock status (bit	1 <sup>st</sup> bit	Interlock		
mask)*	3 <sup>rd</sup> bit	Over current	R	0800
muony	4 <sup>th</sup> bit	Overheat (warning)		
	5 <sup>th</sup> bit External NTC Interlock			
	Lower limit		R/W	0A05
NTC sensor	Upper limit		R/W	0A06
temperature (0.1°)	Measured value		R	0AE4
	B <sub>25/100</sub>		R/W	OB0E
PCB temperature (0.1°)	Measured value		R	0AF4

<sup>\*</sup>If temperature of the device reaches the over temperature warning threshold the overheat flag will be set. If the device is in over temperature protection state, then it will be set overheat and over current flags together.

The maximum duration depends on the set value of the frequency. When you change frequency, a new value of the maximum duration is compute automatically. The duration of pulse cannot be less than 2 ms and more than period of frequency minus 2 ms. For low frequencies the duration cannot be more than 5000 ms.

Set the zero frequency to switch the device into CW mode or set not zero frequency value to switch the device into QCW (long pulses) mode. If you try to set a value more or less than limits, then the value will be rounded to limit. Any attempts to set a new state of the device, except "start", forcibly switch the device to the state "stop". Some states of the device are mutually exclusive, for example, if you set "Ext. Enable", then you will not be able to set the state "start". If you send "start" and "stop" commands to each other, the device will save all parameters in the internal memory. The saving process lasts about 300 ms. In this time the device does not respond to any actions. The device is able to save the next parameters in the internal memory:

- Frequency with limits;
- Duration with limits;
- Current with limits and calibration;
- Temperature limits and B<sub>25/100</sub>;
- Settings of the RS protocol extension (see section «the protocol extension»).

# 15. Digital control description (extended)

**WARNING!** Extended protocol recommend for advanced users only. In addition, it might be used for integration of the device with other devices.

Use the parameter number 0704h for configure the extended protocol. In extended protocol, you can enable and disable the next options: checksum (CRC 8-bit CCITT), return a new value of parameter after P-type commands, change baud-rate, change protocol-mode (text-plain or binary).

The description of the protocol extension command

Action			R\W	HEX-number of parameters	
	0 bit 1 – the device supports this option				
	1st bit	Checksum $(1 - on, 0 - off)$		0704	
Information about the extended protocol	2nd bit	Return a new value for P- type commands (1 – on, 0 – off)			
	3-5 bits	Baud - rate 0 - 2400 1 - 9600 2 - 10417 3 - 19200 4 - 57600 5 - 115200 (default)	R		
	6th bit	Data exchange mode (1 – binary, 0 –text-plain)			
	On checksum (CS)	0002h <sup>1</sup>		0704	
	Off checksum	0004h <sup>1</sup>			
Configuring of the extended protocol	Return a new value for P-type commands	0008h <sup>1</sup>			
	Do not return answer for P-type commands	0010h <sup>2</sup>			
	Set new baud- rate(baud) <sup>3</sup>	0100h - 2400 0120h - 9600 0140h - 10 417 0180h - 57 600 01A0h - 115200	W		
	Text-plain mode on <sup>4</sup>	0200h			
	Binary mode on	0400h			

<sup>&</sup>lt;sup>2</sup> In binary mode the specified commands are ignored by the device

<sup>&</sup>lt;sup>3</sup> Here are binary numbers

<sup>&</sup>lt;sup>4</sup> For more information, see section "binary mode"

#### **Text-plain mode**

All commands in text-plain mode should be in ASCII.

**WARNING!** If you enable the checksum it will change format of commands. After <CR> symbol you will be write 2 bytes of checksum and last byte will be <LF> (0Ah – "new line" symbol). Checksum is computed for all bytes of command before checksum bytes (including<CR> symbol).

All answers of the device will also contain a checksum, including K-type and E-type answers. Checksum is computed by CRC-CCITT-8 algorithm. This is the main difference between the format of commands for the extended protocol and standard protocol.

The format of commands for enabled checksum

Number of byte	Value	Comment
1	P,J,K,E	Type of command
2-5	Number of parameter	Hex-number of value
6	'space' symbol (20h)	(does not use for J and E-type commands)
7-10	New value of the parameter	Hex-value of parameter (does not use for J and E-type commands)
11	'return carriage' symbol <cr> (0Dh)</cr>	End of value
12-13	Checksum	CRC checksum of the first 11 bytes (for J and E-type commands checksum is computed for the first 6 bytes), including <cr> symbol.</cr>
14	'new line' symbol <lf> (0Ah)</lf>	End of command

#### **Possible problems**

- 1. The device waiting for symbol <LF>. If <LF> symbol does not received and buffer is overflowed, then all symbols after overload will be processed as a new command. The device returns an error. In this case, it is recommended to send the <LF> symbol. The device will generate an error and clear the buffer for the next command.
- 2. All symbols after the <LF> symbol will be processed as a new command.

#### **Binary mode**


The binary mode has a significant difference. In this mode, data are exchanged in binary form. Length of any type of command is 8 bytes! In this mode, next options are <u>always enable</u> and you cannot disable it: return a new value of parameter for P-type commands and checksum. The format of binary mode commands is represented in table 8.


#### The format of binary mode commands

Number of byte	Value	Comment
1	Type of command	50h (P - ascii) – Set a new value of parameter 4Ah (J - ascii) – Get a value of parameter 4Bh (K - ascii) – Answer of the device 45h (E - ascii) – Error
2-3	Number of parameter	Hex-number of parameter
4-5	Value of parameter	Hex-value of parameter. This value is <b>STRICTLY REQUIRED</b> in the binary mode. If this value does not make sense, it will be returned as 0000 (K or E-type commands) or you should set it field to any value (for P or J-type commands).
6	'return carriage' symbol <cr> (0Dh)</cr>	End of value
7	Checksum	CRC checksum of the first 11 bytes (for J and E-type commands checksum is computed for the first 6 bytes), including <cr> symbol.</cr>
8	'new line' symbol <lf> (0Ah)</lf>	End of command

# 16. Mechanical dimensions

All dimensions are in millimeters.






# **Test report**

#### Test conditions:

- Test load 60EPU04 diodes 48 in series, unless otherwise specified;
- The output current is measured with Hall-effect current sensor VAC 4646-X201;
- The temperature is measured at the hottest point of PCB and at the Inductor;
- Cooling: conduction to a thermostated heatsink, additional air cooling;
- Supply voltage is 55V.

#### Measuring instruments:

- Oscilloscope Tektronix TDS2002B;
- Arbitrary / function generator Tektronix AFG1022;
- Infrared thermometer ADA TemPro 300.



60EPU04 typical forward voltage drop characteristics

#### Interface test

+/-15V, 5V	
External enable	
External current set	
Internal enable	
Internal current set	
Interlock	
Overheat	
Over current protection	
RS-232, UART	

#### **QCW** mode test

Parameter	Value			
	Set	Measured	Set	Measured
Current, A	5	-	10	-
Frequency, Hz	100		20	
Duration, ms	5		20	
Rise time, us	-		-	
Fall time, us	-		-	

# **Soft-start test**

# of diodes in series	Rise time @ 10A, us
12 (~10V @ 10A)	
24 (~20V @ 10A)	
36 (~30V @ 10A)	
48 (~40V @ 10A)	

# CW mode test

Parameter	Current, A				
	Set	Data from current monitor	Data from RS-232	Measured via 4646- X201 sensor	
	Ext. 2A (1V)				
	Ext. 4A (2V)				
	Ext. 6A (3V)				
	Ext. 8A (4V)				
Value	Ext. 10A (5V)				
	Int. 2A				
	Int. 4A				
	Int. 6A				
	Int. 8A				
	Int. 10A				

# **Continuous test**

Parameter	Current, A	Time, min	Temperature on built-in sensor, °C
Value	10A	0	
		10	
		20	
		30	

Tested by	Date	
Testen nv	Date	